ADD-ON PACKAGES

High Water Alarm Package

Install the High Water Alarm Package in the sterilizer piping as shown on drawing SK47536, but do not install the alarm buzzer box.

Figure 13. INSTALLING LIQUID LEVEL SENSOR